



华标计量检测（深圳）有限公司

HUA BIAO JI LIANG JIAN CE (SHEN ZHEN) YOU XIAN GONG SI

校准证书

CALIBRATION CERTIFICATE



中国认可
国际互认
校准
CALIBRATION
CNAS L12368



第 1 页, 共 3 页
Page of

证书编号

Certificate No.

HB202605207417

委托方

Client

TRIMMING SHOP 007 LTD

委托方地址

Address

UNIT 3, STATION APPROACH, WINDMILL LANE, CHESHUNT, EN8
9A UNITED KINGDOM

仪器名称

Description

M1级标准校正砝码

型号规格

Model/Type

500g

制造厂

Manufacturer

蓬莱鸿鑫平衡器有限公司

出厂编号

Serial No.

/

管理编号

Asset No. /

样品接收日期

Date of Receipt

2026-01-01

校准日期

Date of Calibration

2026-01-01

批准人:

Approved Signatory

程伟才

审核:


Inspected by

吴玄德

校准:

Tested by

宁亮



地址: 广东省深圳市宝安区新玉路48号801室

Add: Room 801, No. 48 Xinyu Road, Bao'an District, Shenzhen City Guangdong Province

电话 (Tel): 0755-2330 4589

邮政编码 (Post Code): 518104

传真 (Fax): 0755-27211132



校准说明

DIRECTIONS OF CALIBRATION

证书编号：HB202605207417

第 2 页，共 3 页

Certificate No.

Page of

1. 本实验室出具的数据均可溯源至国家计量基准。

(All data issued by ZD Test are traced to National Primary Standards.)

2. 本结果仅对当次被测样品有效，如有疑问请在15个工作日反馈。

(The results is ONLY valid for the tested sample, please feedback to us within 15 working days if you have any question.)

3. 本证书编号具有唯一性，若替换证书，自发出后原证书即可作废。

(Each certificate has a unique number. If replaced the certificate, the original certificate will be invalid once the new certificate number is issued.)

4. 证书中如有最大允许误差、判定结果，仅供参考，其中“P”代表“Pass”，“F”代表“Fail”。

(In the datasheet, MPE & determination is only for reference, "P" represents "Pass" and "F" represents "Fail".)

5. 本次校准的技术依据及获认可的能力范围，超出范围的内容未被认可。

(Reference documents and Accredited Scopes for Calibration, Beyond the Scope has not been accredited):

参照JJG 99-2022《砝码检定规程》《Verification regulation for weights》

6. 本次校准使用的主要测量标准 (Main Standards of Measurement Used in the Calibration.):

名称/型号	编号	证书号	有效期至	技术特征
Description/Model	Serial No.	Certificate No.	Due Date	Technique Character
电子天平	19094	20250329156	2026-03-28	II级
砝码	2073	LZM202560065	2026-12-05	M1级

7. 校准地点、环境条件 (Place and environmental conditions of the calibration)

地点	客户现场	温度	22 °C	相对湿度	55%
Place		Temperature		Relative Humidity	

8. 建议复校时间间隔:

12 个月，送检单位可根据实际情况自主决定。

Suggested calibration interval is

12month or it can be altered depending on the actual usage of the user.



校准结果

Result of Calibration

证书编号：HB202605207417

Certificate No.

第 3 页，共 3 页

Page of

1、外观及各部分功能相互作用(Appearance & Interaction of parts): 正常(Normal)

2、示值校准(Calibration indication):

标称质量值 Nominal quality value	实测值约定质量修正值 Agreed quality correction value for measured values	允差 MPE	结论 Conclusion
500g	5mg	±25mg	P

备注(Notes):

1.测量结果的扩展不确定度： $U=0.01\text{mg}\sim 1.2\text{mg}$ ， $k=2$
(Expanded uncertainty of measuring results)

2.依据：JJF 1059.1-2012 《测量不确定度评定与表示》。

In accordance with: JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement.

(以下空白)
(The below is blank)